
	<p style="text-align: center;">استمارة مجال الاعتماد Scope of Accreditation form</p>	<p style="text-align: center;">نظام الاعتماد العراقي IQAS</p>
<p><b>Organization address:</b> Iraq-Kurdistan region- Erbil- main Gazna road</p>	<p><b>Organization name :</b> Modern Surveying for Testing and Calibration Laboratories</p>	<p><b>Accreditation no.:</b> TL 021</p>
<p><b>Signature :</b></p> 	<p style="text-align: center;">Accreditation is valid From 14/05/2020 To 13/05/2022</p>	<p><b>Issue no. :</b> 003</p>

Testing Field	Type of Test	Test Object or Product	Reference to Standardized Method
Mechanical testing	Determination of the Seat tightness of metal and soft-Seated Pressure Safety Valves & Pressure Relief Valves	Pressure Safety Valves & Pressure Relief Valves (PSV, PRV, PVRV and Pilot Valve etc.): -Pneumatic (-0.9 to 300 Bar) -Hydraulic (0.2 to 1200 Bar)	API 527 API 576 MSL/TP/P/02
Electro-Mechanical testing	To Detect and adjust the set point and dead band of the pressure switch	Pressure Switches: -Pneumatic (-0.9 to 200 Bar) -Hydraulic (0 to 1200 Bar)	MSL/TP/P/03