



# PERRY JOHNSON LABORATORY ACCREDITATION, INC.

## Certificate of Accreditation

*Perry Johnson Laboratory Accreditation, Inc. has assessed the Laboratory of:*

***Modern Surveying Calibration & Testing Labs***  
***6, Main Gazna Road, Erbil, Kurdistan, Iraq***

*(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:*

**ISO/IEC 17025:2017**

This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (as outlined by the joint ISO-ILAC-IAF Communiqué dated April 2017):

***Mechanical Testing***  
***(As detailed in the supplement)***

Accreditation claims for such testing and/or calibration services shall only be made from addresses referenced within this certificate. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

Tracy Szerszen  
President

Perry Johnson Laboratory  
Accreditation, Inc. (PJLA)  
755 W. Big Beaver, Suite 1325  
Troy, Michigan 48084

*Initial Accreditation Date:*

September 17, 2020

*Issue Date:*

September 17, 2020

*Expiration Date:*

December 31, 2022

*Accreditation No.:*

106457

*Certificate No.:*

L20-583

*The validity of this certificate is maintained through ongoing assessments based on a continuous accreditation cycle. The validity of this certificate should be confirmed through the PJLA website: [www.pjilabs.com](http://www.pjilabs.com)*



# Certificate of Accreditation: Supplement

## Modern Surveying Calibration & Testing Labs

6, Main Gazna Road, Erbil, Kurdistan, Iraq  
Contact Name: Charanjith PR Phone: 971561188358

Accreditation is granted to the facility to perform the following testing:

FIELD OF TEST	ITEMS, MATERIALS OR PRODUCTS TESTED	SPECIFIC TESTS OR PROPERTIES MEASURED	SPECIFICATION, STANDARD METHOD OR TECHNIQUE USED	RANGE (WHERE APPROPRIATE) AND DETECTION LIMIT
Mechanical <sup>FO</sup>	Safety Valves	Determination of Seat Tightness of metal and soft seated pressure safety valves and pressure relief valves	Procedures: - MSL/TP/P/02 Based on API 576 & API 527 Test Bench for PSV Testing Ventil VC40-SRV & VCB-100 Automated Pressure Calibrator-Additel-ADT761-D Automated Pressure Calibrator Additel-ADT761-HA Standard Pressure Gauges -Crystal -Xp2i	-1 Bar to 1 Bar 0 Bar to 7 Bar 0 Bar to 30 Bar 0 Bar to 200 Bar 0 Bar to 700 Bar

1. The presence of a superscript FO means that the laboratory performs testing of the indicated parameter both at its fixed location and onsite at customer locations. Example: Outside Micrometer<sup>FO</sup> would mean that the laboratory performs this testing at its fixed location and onsite at customer locations.