
**PERRY JOHNSON
LABORATORY ACCREDITATION,
INC.**

Certificate of Accreditation

Perry Johnson Laboratory Accreditation, Inc. has assessed the Laboratory of:

Modern Surveying Calibration & Testing Labs
6, Main Gazna Road, Erbil, Kurdistan Iraq

*(Hereinafter called the Organization) and hereby declares that Organization is accredited
in accordance with the recognized International Standard:*

ISO/IEC 17025:2017

This accreditation demonstrates technical competence for a defined scope and the
operation of a laboratory quality management system
(as outlined by the joint ISO-ILAC-IAF Communiqué dated April 2017):

***Mass, Force & Weighing, Time & Frequency, Dimensional,
Mechanical, Electrical, Thermodynamic and Chemical Devices***
(As detailed in the supplement)

Accreditation claims for such testing and/or calibration services shall only be made from addresses referenced within this certificate. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:



Tracy Szerszen
President
Perry Johnson Laboratory
Accreditation, Inc. (PJLA)
755 W. Big Beaver, Suite
1325Troy, Michigan 48084

Initial Accreditation Date:

September 17, 2020

Issue Date:

November 11, 2022

Expiration Date:

December 31, 2024

Accreditation No.:

106457

Certificate No.:

L22-797




Certificate of Accreditation: Supplement

ISO/IEC 17025:2005 and DoD-ELAP

Laboratory Name
8100 Address Drive, City, ST 99999

Accreditation is granted to the facility to perform the following testing:

The validity of this certificate is maintained through ongoing assessments based on a continuous accreditation cycle. The validity of this certificate should be confirmed through the PJLA website: www.pjlab.com





Certificate of Accreditation: Supplement

Modern Surveying Calibration & Testing Labs

6, Main Gazna Road, Erbil, Kurdistan, Iraq
 Contact Name: Charanjith PR Phone: +964 7517596829

Accreditation is granted to the facility to perform the following calibrations:

Mechanical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Pressure Gauge Differential Pressure Indicator Pressure Sensors Pressure Transducer Pressure Transmitter ^{FO}	Up to 1 000 mBar	0.001 %	Automated Pressure Calibrator-Additel- ADT761-D Standard Pressure Gauge/ Modules Fluke & Additel MSL/CP/P/04 Based on DKD-R-6-1
	1 Bar to 40 Bar	0.002 5 %	Automated Pressure Calibrator Additel- ADT761-HA Standard Pressure Gauge/ Modules Fluke & Additel MSL/CP/P/04 Based on DKD-R-6-1
	40 Bar to 200 Bar	0.005 %	Pressure Balance DH- Budenberg - CPB 5800 Standard Pressure Gauge/ Modules Fluke & Additel MSL/CP/P/04 Based on DKD-R-6-1
Pressure Gauge Differential Pressure Indicator Pressure Sensors Pressure Transducer Pressure Transmitter ^{FO}	200 Bar to 1 200 bar	0.006 %	Pressure Balance DH- Budenberg - CPB 5800 Standard Pressure Gauge/ Modules Fluke & Additel MSL/CP/P/04 Based on DKD-R-6-1
Vacuum Gauges Vacuum Transducer Vacuum Transmitters Vacuum Sensors ^{FO}	-900 mBar to -0.00 mBar	0.001 %	Automated Pressure Calibrator-Additel- ADT761-D Standard Pressure Gauge/ Modules Fluke & Additel MSL/CP/P/04 Based on DKD-R-6-1



Certificate of Accreditation: Supplement

Modern Surveying Calibration & Testing Labs

6, Main Gazna Road, Erbil, Kurdistan, Iraq
 Contact Name: Charanjith PR Phone: +964 7517596829

Accreditation is granted to the facility to perform the following calibrations:

Thermodynamic

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Dry Block Calibrators ^F	-40 °C to 140 °C	0.039 °C	Fluke Black Stack 1560 with modules 2560,2561,2562,2565 & 2566, SPRT Fluke 5698 PRT's Fluke-5626 & Fluke- 5628, EURAMET Calibration Guide No.13
	140 °C to 300 °C	0.078 °C	
	300 °C to 400 °C	0.12 °C	
	400 °C to 650 °C	0.16 °C	
Infrared Thermometers: ^F	-30 °C	2 °C	Portable Infrared Calibrator- Fluke 9133 ASTM E2847
	-25 °C	1.8 °C	
	-20 °C	1.6 °C	
	-10 °C	1.1 °C	
	-5 °C	0.94 °C	
	0 °C	0.72 °C	
	23 °C	0.31 °C	
	50 °C	0.67 °C	
	75 °C	1.1 °C	
	100 °C	1.5 °C	
	125 °C	1.9 °C	
	150 °C	2.3 °C	
Temperature Chamber ^{FO} System Accuracy Test	-40 °C to -18 °C	0.58 °C	Fluke-2638A Hydra Series III Data Acquisition System DATA Logger Rotronics-HL-20D High Temperature Data Logger- Madgetech-Hi Temp 140 Thermocouple Wire- PRT Sensor BS EN 60068-3-5
	-18 °C to 0 °C	0.58 °C	
	0 °C to 8 °C	0.58 °C	
	8 °C to 45 °C	0.1 °C	
	45 °C to 100 °C	0.1 °C	
	100 °C to 250 °C	0.13 °C	
Temperature Chamber ^{FO} Thermal Uniformity Survey	-40 °C to -18 °C	0.96 °C	
	-18 °C to 0 °C	2.1 °C	
	0 °C to 8 °C	2.1 °C	
	8 °C to 45 °C	0.14 °C	
	45 °C to 100 °C	0.21 °C	
	100 °C to 250 °C	0.32 °C	



Certificate of Accreditation: Supplement

Modern Surveying Calibration & Testing Labs

6, Main Gazna Road, Erbil, Kurdistan, Iraq
 Contact Name: Charanjith PR Phone: +964 7517596829

Accreditation is granted to the facility to perform the following calibrations:

Thermodynamic

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Digital Thermometer with RTD and/or Thermocouple ^{FO}	-40 °C to 0 °C	0.046 °C	PRT Fluke 5626 & 5628, Black Stack Thermometer Read Out – Fluke-1560 SPRT Module Fluke- 2560. High -Temp PRT Module Fluke-2561 Precision Thermocouple Module -Fluke- 2565 Precision Baths: Fluke 6331, 7381, Multi-function Calibrator Wika- CTM9100-150 Metrology Wells: Fluke 9170, 9173 Type S Thermocouple Standard-Fluke 5650 MSL/CP/T/05
	0 °C to 50 °C	0.046 °C	
	50 °C to 100 °C	0.057 °C	
	100 °C to 175 °C	0.074 °C	
	175 °C to 250 °C	0.076 °C	
	250 °C to 400 °C	0.16 °C	
	400 °C to 600 °C	0.2 °C	

Dimensional

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Outside Micrometer ^F	Up to 25 mm Resolution: 0.001 mm	(2.6 + 0.12 L) μ m	Gauge Block Set-Mitutoyo- 516-106-10 Gauge Block Set-Tesa BS:870
	Up to 25 mm Resolution: 0.01 mm	(5.8+0.05 L) μ m	
Inside Micrometer ^F	Up to 25 mm Resolution: 0.01 mm	(6.5 + 0.15 L) μ m	Gauge Block Set-Mitutoyo- 516-106-10 Gauge Block Set-Tesa BS:959
Depth Micrometer ^F	Up to 25 mm Resolution: 0.01 mm	(5.8 + 0.01 L) μ m	Gauge Block Set-Tesa BS:6468
Calipers (Vernier, Dial & Digital) ^{FO}	Up to 600 mm Resolution 0.01 mm	(6.1 + 0.3 L) μ m	Caliper checker Mitutoyo- 515-556-2 BS:887
	Up to 300 mm Resolution 0.02 mm	(11 + 0.22 L) μ m	
	Up to 600 mm Resolution 0.05 mm	(28 + 0.08 L) μ m	
Measuring Tapes ^F	Up to 30 m	(580 + 0.12 L) μ m	Measuring Scale & Tape Calibration System Octagon MSTC-1000 OIML: R 35-1 OIML: R 35-2
Steel Ruler ^F	Up to 1000 mm	580 μ m	
Indicator (Dial/Digital) ^F	Up to 100 mm	6 μ m	Dial Indication Tester- Mitutoyo 170- 102-12 Gauge Block Set-Tesa MSL/CP/D/07 Based on BS EN ISO 463
Ultrasonic Thickness Gauge ^F	2.5 mm to 20 mm	10 μ m	Five Step Block ASTM-E317



Certificate of Accreditation: Supplement

Modern Surveying Calibration & Testing Labs

6, Main Gazna Road, Erbil, Kurdistan, Iraq
 Contact Name: Charanjith PR Phone: +964 7517596829

Accreditation is granted to the facility to perform the following calibrations:

Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Equipment to Measure DC Voltage ^{FO}	Up to 330 mV	25 μ V/V + 1 μ V	Multi-Product Calibrator - Fluke 5522A MSL/CP/E/01
	0.33 V to 3.3 V	14 μ V/V + 2 μ V	
	3.3 V to 30 V	9 μ V/V + 20 μ V	
	30 V to 330 V	12 μ V/V + 150 μ V	
	330 V to 1000 V	12 μ V/V + 1.5 mV	
Equipment to Measure AC Voltage ^{FO}	Up to 33 mV @ 45 Hz to 1 kHz	600 μ V/V + 6 μ V	Multi-Product Calibrator - Fluke 5522A MSL/CP/E/01
	33 mV to 330 mV @ 45 Hz to 1 kHz	120 μ V/V + 8 μ V	
	0.33 V to 3.3 V @ 45 Hz to 1 kHz	82 μ V/V + 160 μ V	
	3.3 V to 33 V @ 45 Hz to 1 kHz	55 μ V/V + 600 μ V	
	33 V to 330 V @ 45 Hz to 1 kHz	65 μ V/V + 2 000 μ V	
	330 V to 1000 V @ 45 Hz to 1 kHz	90 μ V/V + 10 mV	
Equipment to Measure DC Current ^{FO}	100 μ A to 330 μ A	4 μ A/A + 0.02 μ A	Multi-Product Calibrator - Fluke 5522A MSL/CP/E/01
	0.33 mA to 3.3 mA	25 μ A/A + 0.05 μ A	
	3.3 mA to 33 mA	25 μ A/A + 0.25 μ A	
	33 mA to 330 mA	25 μ A/A + 2.5 μ A	
	0.33 A to 1.1 A	41 μ A/A + 40 μ A	
	1.1 A to 3 A	50 μ A/A + 40 μ A	
	3 A to 11 A	470 μ A/A + 750 μ A	
	11 A to 20 A	800 μ A/A + 1500 μ A	
Clamp - On Meters to Measure DC Current ^{FO}	10 A to 16.5 A	0.19 % + 1.6 mA	Multi-Product Calibrator - Fluke 5522A Fluke 5500A (Coil) MSL/CP/E/01
	16.5 A to 150 A	0.19 % + 12 mA	
	150 A to 1 000 A	0.19 % + 39 mA	
Clamp - On Meters to Measure AC Current ^{FO}	10 A to 16.5 A @ 45 Hz to 1 kHz	0.22 % + 2.33 mA	Multi-Product Calibrator - Fluke 5522A Fluke 5500A (Coil) MSL/CP/E/01
	16.5 A to 150 A @ 45 Hz to 1 kHz	0.22 % + 19.38 mA	
	150 A to 1 000 A @ 45Hz to 1 kHz	0.22 % + 69.77 mA	



Certificate of Accreditation: Supplement

Modern Surveying Calibration & Testing Labs

6, Main Gazna Road, Erbil, Kurdistan, Iraq
 Contact Name: Charanjith PR Phone: +964 7517596829

Accreditation is granted to the facility to perform the following calibrations:

Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Equipment to Measure AC Current ^{FO}	29 μ A to 330 μ A @ 45 Hz to 1 kHz	0.12 % + 0.1 μ A	Multi-Product Calibrator - Fluke 5522A MSL/CP/E/01
	0.33 mA to 3.3 mA @ 45Hz to 1 kHz	0.11 % + 0.15 μ A	
	3.3 mA to 33 mA @ 45 Hz to 1 kHz	0.048 % + 2 μ A	
	33 mA to 330 mA @ 45 Hz to 1 kHz	0.05 % + 20 μ A	
	0.33 A to 1.1 A @ 45 Hz to 1 kHz	0.06 % + 0.1 mA	
	1.1 A to 3 A @ 45 Hz to 1 kHz	0.08 + 0.1 mA	
	3 A to 11 A @ 45 Hz to 1 kHz	0.09 % + 2 mA	
	11 A to 20 A @ 45 Hz to 1 kHz	0.2 % + 5 mA	
Equipment to Measure Capacitance ^F	0.19 nF to 0.3999 nF	0.61 % + 0.01 nF	Multi-Product Calibrator - Fluke 5522A MSL/CP/E/01
	0.4 nF to 1.099 9 nF	0.58 % + 0.01 nF	
	1.1 nF to 3.299 9 nF	0.58 % + 0.01 nF	
	3.3 nF to 10.999 9 nF	0.30 % + 0.01 nF	
	11 nF to 32.999 9 nF	0.30 % + 0.1 nF	
	33 nF to 109.999 nF	0.30 % + 0.1 nF	
	110 nF to 329.999 nF	0.30 % + 0.3 nF	
	0.33 μ F to 1.0999 9 μ F	0.30 % + 1 nF	
	1.1 μ F to 3.299 99 μ F	0.30 % + 3 nF	
	3.3 μ F to 10.999 9 μ F	0.30 % + 10 nF	
	11 μ F to 32.999 9 μ F	0.47 % + 30 nF	
	33 μ F to 109.999 μ F	0.54 % + 100 nF	
	110 μ F to 329.999 μ F	0.52 % + 300 nF	
	0.33 mF to 1.099 99 mF	0.52 % + 1 μ F	
	1.1 mF to 3.299 99 mF	0.52 % + 3 μ F	
	3.3 mF to 10.999 9 mF	0.52 % + 10 μ F	
11 mF to 32.999 9 mF	0.87 % + 30 μ F		
33 mF to 110 mF	1.3 % + 100 μ F		



Certificate of Accreditation: Supplement

Modern Surveying Calibration & Testing Labs

6, Main Gazna Road, Erbil, Kurdistan, Iraq
 Contact Name: Charanjith PR Phone: 971561188358

Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED		
Equipment to Measure Resistance (Fixed Points) ^F	1 m Ω	0.000 1 m Ω	Burster- 1240 Calibration Resistor		
	10 m Ω	0.001 m Ω			
	100 m Ω	0.01 m Ω			
	Equipment to Measure Resistance ^F	1 Ω	0.000 011 Ω	Fluke 742 A Resistance Standards	
		10 Ω	0.001 Ω		
		100 Ω	0.01 Ω		
		Equipment to Measure Resistance ^F	1 K Ω	0.000 08 K Ω	MSL/CP/E/01
			10 K Ω	0.000 8 K Ω	
			1 M Ω	0.008 K Ω	
10 M Ω			0.002 M Ω		
0.1 Ω to 11 Ω			0.006 % + 0.5 m Ω	Multi-Product Calibrator - Fluke 5522A	
11 Ω to 33 Ω			0.001 % + 1 m Ω		
33 Ω to 110 Ω	0.001 % + 4 m Ω				
110 Ω to 330 Ω	0.001 % + 4 m Ω				
0.33 K Ω to 1.1 K Ω	0.001 % + 4 m Ω				
1.1 K Ω to 3.3 K Ω	0.001 % + 4 m Ω				
3.3 K Ω to 11 K Ω	0.001 % + 1 Ω				
11 K Ω to 33 K Ω	0.08 % + 4 Ω				
33 K Ω to 110 K Ω	0.001 % + 4 Ω				
110 K Ω to 330 K Ω	0.001 % + 11 Ω				
0.33 M Ω to 1.1 M Ω	0.001 % + 36 Ω				
1.1 M Ω to 3.3 M Ω	0.006 % + 0.2 k Ω				
3.3 M Ω to 11 M Ω	0.036 % + 1.5 k Ω				
11 M Ω to 33 M Ω	0.1 % + 8.3 k Ω				
33 M Ω to 110 M Ω	0.01 % + 55 k Ω				
110 M Ω to 330 M Ω	0.01 % + 0.1 M Ω				
330 M Ω to 1 100 M Ω	0.74 % + 0.1 M Ω				
Equipment to Measure Inductance ^F @ 1 kHz	10 μ H to 10 mH	1 % + 1.5 μ H	Programmable Inductance Substituter IET Labs- PLS -1492		
	10 mH to 100 mH	1 % + 1.5 μ H			
	100 mH to 1 H	2 % + 1.5 μ H			
	1 H to 10 H	2 % + 1.5 μ H	MSL/CP/E/01		



Certificate of Accreditation: Supplement

Modern Surveying Calibration & Testing Labs

6, Main Gazna Road, Erbil, Kurdistan, Iraq
 Contact Name: Charanjith PR Phone: +964 7517596829

Accreditation is granted to the facility to perform the following calibrations:

Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Equipment to Output AC Current ^F	1 μ A to 200 μ A @ 45 Hz to 1 KHz	0.054 % + 20 nA	Reference Multimeter Fluke -8508A
	0.2 mA to 2 mA @ 45 Hz to 1 KHz	0.032 % + 0.2 μ A	MSL/CP/E/01
	2 mA to 20 mA @ 45 Hz to 1 KHz	0.033 % + 2 μ A	
	20 mA to 200 mA @ 45 Hz to 1 KHz	0.031 % + 20 μ A	
	0.2 A to 2 A @ 45 Hz to 1 KHz	0.063 % + 0.2 mA	
	2 A to 20 A @ 45 Hz to 1 KHz	0.84 % + 2 mA	
Equipment to Output DC Current ^F	1 μ A to 200 μ A	42 μ A/A + 0.023 μ A	Reference Multimeter Fluke -8508A
	0.2 mA to 2 mA	26 μ A/A + 0.003 μ A	MSL/CP/E/01
	2 mA to 20 mA	28 μ A/A + 0.03 μ A	
	20 mA to 200 mA	28 μ A/A + 0.03 μ A	
	0.2 A to 2 A	220 μ A/A + 0.02 μ A	
	2 A to 20 A	0.48 mA/A + 0.4 mA	
Equipment to Output AC Voltage ^F	1 mv to 200 mv @ 45 Hz to 1 kHz	0.011 % + 0.002 mV	Reference Multimeter Fluke -8508A
	0.2 V to 2 V @ 45 Hz to 1 kHz	0.0086 % + 20 μ V	MSL/CP/E/01
	2 V to 20 V @ 45 Hz to 1 kHz	0.0086 % + 0.2 mV	
	20 V to 200 V @ 45 Hz to 1 kHz	0.009 % + 2 mV	
	200 V to 1 000 V @ 45 Hz to 1kHz	0.011 % + 2 mV	
Equipment to Output DC Voltage ^F	1 mv to 200 mv	10 μ V/V + 0.07 μ V	Reference Multimeter Fluke -8508A
	0.2 V to 2 V	5 μ V/V + 1.4 μ V	MSL/CP/E/01
	2 V to 20 V	5 μ V/V + 40 μ V	
	20 V to 200 V	8 μ V/V + 36 μ V	
	200 V to 1 000 V	10 μ V/V + 0.49 mV	
Equipment to Output Inductance ^F	100 μ H to 1 mH	0.12 %	Precision RLC Digibridge
	1 mH to 1 H	0.035 %	IET Labs- 1693
	1 H to 5 H	0.014 %	
	5 H to 10 H	0.25 %	MSL/CP/E/01



Certificate of Accreditation: Supplement

Modern Surveying Calibration & Testing Labs

6, Main Gazna Road, Erbil, Kurdistan, Iraq
 Contact Name: Charanjith PR Phone: +964 7517596829

Accreditation is granted to the facility to perform the following calibrations:

Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Equipment to Output Resistance ^F	1 m Ω to 2 Ω	0.002 % Rdg	Reference Multimeter Fluke -8508A MSL/CP/E/01
	2 Ω to 20 Ω	0.002 % Rdg	
	20 Ω to 200 Ω	0.000 9 % Rdg	
	0.2 K Ω to 2 K Ω	0.001 % Rdg	
	2 k Ω to 20 K Ω	0.001 % Rdg	
	20 K Ω to 200 K Ω	0.001 % Rdg	
	0.2 M Ω to 2 M Ω	0.001 % Rdg	
	2 M Ω to 20 M Ω	0.003 % Rdg	
	20 M Ω to 200 M Ω	0.009 % Rdg	
	0.2 G Ω to 2 G Ω	0.1 % Rdg	
Equipment to Output Capacitance ^F @ 1 kHz	1 pF to 10 pF	0.37 % + 0.002 pF	Precision RLC Digibridge ET Labs- 1693 MSL/CP/E/01
	10 pF to 100 pF	0.014 % + 0.003 pF	
	100 pF to 1 000 pF	0.007 % + 0.01 pF	
	1 nF to 10 nF @ 1 kHz	0.010 % + 0.08 pF	
	10 nF to 100 nF @ 1 kHz	0.016 % Rdg	
	100 nF to 1 000 nF	0.01 % Rdg	
	1 μ F to 10 μ F @ 1 kHz	0.01 % Rdg	
	10 μ F to 100 μ F @ 1 kHz	0.008 % Rdg	
	100 μ F to 1000 μ F	0.02 % Rdg	
	1 000 μ F to 10 000 μ F	0.10 % Rdg	
Calibration of Temperature Indicators and Simulators by Electrical Simulation and Measurement (Type J) ^F	-210 $^{\circ}$ C to -100 $^{\circ}$ C	0.21 $^{\circ}$ C	Multi-Product Calibrator - Fluke 5522A EURAMET cg-11
	-100 $^{\circ}$ C to -30 $^{\circ}$ C	0.12 $^{\circ}$ C	
	-30 $^{\circ}$ C to 150 $^{\circ}$ C	0.08 $^{\circ}$ C	
	150 $^{\circ}$ C to 760 $^{\circ}$ C	0.12 $^{\circ}$ C	
	760 $^{\circ}$ C to 1 200 $^{\circ}$ C	0.18 $^{\circ}$ C	
Calibration of Temperature Indicators and Simulators by Electrical Simulation and Measurement (Type K) ^F	-200 $^{\circ}$ C to -100 $^{\circ}$ C	0.25 $^{\circ}$ C	Multi-Product Calibrator - Fluke 5522A EURAMET cg-11
	-100 $^{\circ}$ C to -25 $^{\circ}$ C	0.12 $^{\circ}$ C	
	-25 $^{\circ}$ C to 120 $^{\circ}$ C	0.09 $^{\circ}$ C	
	120 $^{\circ}$ C to 1 000 $^{\circ}$ C	0.19 $^{\circ}$ C	
	1 000 $^{\circ}$ C to 1 372 $^{\circ}$ C	0.57 $^{\circ}$ C	
Calibration of Temperature Indicators and Simulators by Electrical Simulation and Measurement (Type R) ^F	0 $^{\circ}$ C to 250 $^{\circ}$ C	0.21 $^{\circ}$ C	Multi-Product Calibrator - Fluke 5522A EURAMET cg-11
	250 $^{\circ}$ C to 400 $^{\circ}$ C	0.27 $^{\circ}$ C	
	400 $^{\circ}$ C to 1 000 $^{\circ}$ C	0.35 $^{\circ}$ C	
	1 000 $^{\circ}$ C to 1 767 $^{\circ}$ C	0.47 $^{\circ}$ C	



Certificate of Accreditation: Supplement

Modern Surveying Calibration & Testing Labs

6, Main Gazna Road, Erbil, Kurdistan, Iraq
 Contact Name: Charanjith PR Phone: 971561188358

Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Calibration of Temperature Indicators and Simulators by Electrical Simulation and Measurement (Type S) ^F	0 °C to 250 °C	0.38 °C	Multi-Product Calibrator - Fluke 5522A
	250 °C to 1 000 °C	0.29 °C	
	1 000 °C to 1 400 °C	0.41 °C	EURAMET cg-11
	1 400 °C to 1 767 °C	0.49 °C	

Time & Frequency

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Timer, Stopwatches & Chart Speed ^{FO}	10 s to 36 000 s	0.04 s	Timer/Counter/Analyzer: 300MHz-Tektronix-FCA3100 NIST 960-12 Special Publication

Mass, Force and Weighing Device

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Vertical Cylindrical Tank ^O	100 m ³ to 50 000 m ³	0.1 % of volume	Total station Measuring Tape Dipping Tape Ultrasonic Thickness Gauge ISO 7507-1 & ISO 7507-2



Certificate of Accreditation: Supplement

Modern Surveying Calibration & Testing Labs

6, Main Gazna Road, Erbil, Kurdistan, Iraq
 Contact Name: Charanjith PR Phone: +964 7517596829

Accreditation is granted to the facility to perform the following calibrations:

Chemical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Density Meter ^{FO}	0.65 g/cm ³ to 1.8 g/cm ³	0.000 1 g/cm ³	Density Standard Solution MSL-CP-A-02

1. The CMC (Calibration and Measurement Capability) stated for calibrations included on this scope of accreditation represents the smallest measurement uncertainty attainable by the laboratory when performing a more or less routine calibration of a nearly ideal device under nearly ideal conditions. It is typically expressed at a confidence level of 95 % using a coverage factor k (usually equal to 2). The actual measurement uncertainty associated with a specific calibration performed by the laboratory will typically be larger than the CMC for the same calibration since capability and performance of the device being calibrated and the conditions related to the calibration may reasonably be expected to deviate from ideal to some degree.
2. The laboratories range of calibration capability for all disciplines for which they are accredited is the interval from the smallest calibrated standard to the largest calibrated standard used in performing the calibration. The low end of this range must be an attainable value for which the laboratory has or has access to the standard referenced. Verification of an indicated value of zero in the absence of a standard is common practice in the procedure for many calibrations but by its definition it does not constitute calibration of zero capacity.
3. The presence of a superscript F means that the laboratory performs calibration of the indicated parameter at its fixed location. Example: Outside Micrometer^F would mean that the laboratory performs this calibration at its fixed location.
4. The presence of a superscript FO means that the laboratory performs calibration of the indicated parameter both at its fixed location and onsite at customer locations. Example: Outside Micrometer^{FO} would mean that the laboratory performs this calibration at its fixed location and onsite at customer locations.
5. Measurement uncertainties obtained for calibrations performed at customer sites can be expected to be larger than the measurement uncertainties obtained at the laboratories fixed location for similar calibrations. This is due to the effects of transportation of the standards and equipment and upon environmental conditions at the customer site which are typically not controlled as closely as at the laboratories fixed location.
6. The term L represents length in meter.
7. The term T represents temperature in °C or °F as appropriate to the uncertainty statement.