



PERRY JOHNSON LABORATORY ACCREDITATION, INC.

Certificate of Accreditation

Perry Johnson Laboratory Accreditation, Inc. has assessed the Laboratory of:

Modern Surveying Calibration & Testing Labs
6, Main Gazna Road, Erbil, Kurdistan, Iraq

(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:

ISO/IEC 17025:2017

This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (as outlined by the joint ISO-ILAC-IAF Communiqué dated April 2017):

Mechanical Testing
(As detailed in the supplement)

Accreditation claims for such testing and/or calibration services shall only be made from addresses referenced within this certificate. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

Tracy Szerszen
President

Initial Accreditation Date:

September 17, 2020

Issue Date:

November 11, 2022

Expiration Date:

December 31, 2024

Revision Date:

February 07, 2024

Accreditation No.:

106457

Certificate No.:

L22-798-R1

Perry Johnson Laboratory
Accreditation, Inc. (PJLA)
755 W. Big Beaver, Suite 1325
Troy, Michigan 48084

The validity of this certificate is maintained through ongoing assessments based on a continuous accreditation cycle. The validity of this certificate should be confirmed through the PJLA website: www.pjilabs.com



Certificate of Accreditation: Supplement

Modern Surveying Calibration & Testing Labs

6, Main Gazna Road, Erbil, Kurdistan, Iraq

Contact Name: Charanjith PR Phone: +964-751-759-6829

Accreditation is granted to the facility to perform the following testing:

FIELD OF TEST	ITEMS, MATERIALS OR PRODUCTS TESTED	SPECIFIC TESTS OR PROPERTIES MEASURED	SPECIFICATION, STANDARD METHOD OR TECHNIQUE USED	RANGE (WHERE APPROPRIATE) AND DETECTION LIMIT
Mechanical ^{FO}	Safety Valves	Determination of Seat Tightness of metal and soft seated pressure safety valves and pressure relief valves	Procedures: - MSL/TP/P/02 Based on API 576 & API 527 Test Bench for PSV Testing Ventil VC40-SRV & VCB-100 Automated Pressure Calibrator-Additel-ADT761-D Automated Pressure Calibrator Additel-ADT761-HA Standard Pressure Gauges -Crystal -Xp2i	-1 Bar to 1 Bar Up to 7 Bar Up to 30 Bar Up to 200 Bar Up to 700 Bar

1. The presence of a superscript FO means that the laboratory performs testing of the indicated parameter both at its fixed location and onsite at customer locations.